



STANDARDS
MALAYSIA

Certificate of Accreditation

No: SMM 176

Accredited since: 17 September 1999

This is to certify that

SIRIM CALIBRATION SDN. BHD.
SENAI, JOHOR
MALAYSIA



Scan this QR Code or visit
www.jsm.gov.my/cab-directories
for the current scope of accreditation

has been granted accreditation in respect of the scope of accreditation described in the schedule, subject to the terms and conditions governing the *Skim Akreditasi Makmal Malaysia* (SMM), the Laboratory Accreditation Scheme of Malaysia.

Laboratories accredited under SMM meet the requirements of MS ISO/IEC 17025. This Malaysian Standard is identical with ISO/IEC 17025 published by the International Organization for Standardization (ISO).



(SHAHARUL SADRI BIN ALWI)
Director General
Department of Standards Malaysia

Date of issue: 3 January 2023
(Issue 3, 3 January 2023 replacement of SMM 176
dated 6 December 2017)

NO: SMM 176

Page: 1 of 28

LABORATORY LOCATION:
(PERMANENT LABORATORY)

SIRIM CALIBRATION SDN. BHD.
BANGUNAN SIRIM BERHAD
NO. 3, JALAN TEKNOLOGI 5
TAMAN TEKNOLOGI JOHOR
81400 SENAI, JOHOR
MALAYSIA

FIELDS OF CALIBRATION:

ELECTRICAL, DIMENSIONAL, TEMPERATURE,
MASS, FORCE AND PRESSURE

This laboratory has demonstrated its technical competence to operate in accordance with MS ISO/IEC 17025:2017 (ISO/IEC 17025:2017).

This laboratory's fulfillment of the requirements of ISO/IEC 17025 means the laboratory meets both the technical competence requirements and management system requirements that are necessary for it to consistently deliver technically valid test results and calibrations. The management system requirements in ISO/IEC 17025 are written in language relevant to laboratory operations and operate generally in accordance with the principles of ISO 9001 (see Joint ISO-ILAC-IAF Communiqué dated April 2017).

* The expanded uncertainties are based on an estimated confidence probability of approximately 95% and have a coverage factor of $k=2$ unless stated otherwise.

SCOPE OF CALIBRATION: ELECTRICAL

| Instrument Calibrated/ Measurement Parameter | Range | Calibration and Measurement Capability Expressed as an Uncertainty (\pm)* | Remarks |
|--|--|--|--|
| 1. Measuring Instruments | | | |
| (a) DC Voltage | 0 mV to 330 mV 330 mV to 3.3 V 3.3V to 33 V 33 V to 330 V 330 V to 1020 V | 1.6 μ V 3.2 μ V 20 μ V 180 μ V 1.7 mV | Generation using calibrator model Fluke 5522 A |
| (b) Resistance | 0 Ω to 11 Ω 11 Ω to 33 Ω 33 Ω to 110 Ω 110 Ω to 330 Ω 330 Ω to 1.1k Ω 1.1 k Ω to 3.3 k Ω 3.3 k Ω to 11 k Ω 11 k Ω to 33 k Ω 33 k Ω to 110 k Ω 110 k Ω to 330 k Ω 330 k Ω to 1.1 M Ω 1.1 M Ω to 3.3 M Ω 3.3 M Ω to 11 M Ω 11 M Ω to 33 M Ω 33 M Ω to 100 M Ω 100 M Ω to 330 M Ω 330 M Ω to 1100 M Ω | 220 μ Ω 200 μ Ω 740 μ Ω 1.3 m Ω 4.3 m Ω 12 m Ω 54 m Ω 160 m Ω 450 m Ω 1.7 Ω 9 Ω 34 Ω 140 Ω 1.1 k Ω 2.7 k Ω 79 k Ω 560 k Ω | Generation using calibrator model Fluke 5522 A |

SCOPE OF CALIBRATION: ELECTRICAL

| Instrument Calibrated / Measurement Parameter | Range | Calibration and Measurement Capability Expressed as an Uncertainty (\pm)* | Remarks |
|---|--|---|--|
| 1. Measuring Instruments (continued) | | | |
| (b) Resistance | 0 Ω to 1 Ω 1 Ω to 100 Ω 100 Ω to 10 k Ω 10 k Ω to 100 k Ω 10 k Ω to 10 M Ω | 6 $\mu\Omega$ 1 m Ω 1 m Ω 700 m Ω 22 Ω | Generation using Decade resistance box Yokogawa 2793 & Standard Resistor Yokogawa 2792 |
| (c) DC Current | 0 μ A to 330 μ A 330 mA to 3.3 mA 3.3 mA to 33 mA 33 mA to 330 mA 330 mA to 1.1 mA 1.1 A to 3 A 3 A to 10 A 10 A to 21 A | 5.2 nA 34 nA 910 nA 8.7 μ A 42 μ A 100 μ A 1.4 mA 9 mA | Generation using calibrator model Fluke 5522 A |
| (d) AC Voltage (See Matrix A) | See Matrix A | See Matrix A | Generation using calibrator model Fluke 5522 A |

Matrix A
AC Voltage Measurement

| Range | Frequency | | | | | | | | | |
|-----------------|----------------------|----------------------|----------------------|----------------------|-----------------------|-----------------------|------------------------|------------------------|-------------------------|--------------------------|
| | 10 Hz to 45 Hz | 45 Hz to 1 kHz | 1 kHz to 5 kHz | 5kHz to 10 kHz | 1 kHz to 10 kHz | 45 Hz to 10 kHz | 10 kHz to 20 kHz | 20 kHz to 50 kHz | 50 kHz to 100 kHz | 100 kHz to 500 kHz |
| 0 to 33 mV | 2 μ V | - | - | - | - | 2.4 μ V | 2 μ V | 3.0 μ V | 5.0 μ V | 26 μ V |
| 33 mV to 330 mV | 13 μ V | - | - | - | - | 7.3 μ V | 8 μ V | 9.5 μ V | 12 μ V | 20 μ V |
| 330 mV to 3.3 V | 73 μ V | - | - | - | - | 59 μ V | 71 μ V | 100 μ V | 140 μ V | 3 mV |
| 3.3 V to 33 V | 810 μ V | - | - | - | - | 710 μ V | 810 μ V | 1.8 mV | 2.8 mV | - |
| 33 V to 330 V | - | 9.9 mV | - | - | 8.2 mV | - | 10 mV | 22 mV | 74 mV | - |
| 330 V to 1020 V | - | 33 mV | 35 mV | 37 mV | - | - | - | - | - | - |

Scan this QR Code or visit www.ism.gov.my/cab-directories for the current scope of accreditation

NO: SAMM 176

Page: 3 of 28

SCOPE OF CALIBRATION: ELECTRICAL

| Instrument Calibrated / Measurement Parameter | Range | Calibration and Measurement Capability Expressed as an Uncertainty (\pm)* | Remarks |
|---|---|---|---|
| 1. Measuring Instruments (continued) | | | |
| (e) AC Current (See Matrix B) | See Matrix B | See Matrix B | Generation using calibrator model Fluke 5522 A |
| (f) Frequency Measure (Measuring Instrument) | DC 0 MHz to 60 MHz 60 MHz to 2 GHz | 5 Hz 1 Hz | Generate using Function Generator HP 3325 B / Signal Generator Agilent E4433B |
| (g) Level | + 7 dBm to -70 dBm | 1.5 dBm | |
| (h) Capacitance | 1 μ F to 3.3 μ F 3.3 μ F to 11 μ F 11 μ F to 33 μ F 33 μ F to 110 μ F 110 μ F to 330 μ F 0.33 mF to 1.1 mF 1.1 mF to 3.3 mF 3.3 mF to 11 mF 11 mF to 33 mF 33 mF to 110 mF | 9.3 nF 21 nF 85 nF 290 nF 920 nF 6.4 μ F 19 μ F 37 μ F 130 μ F 540 μ F | Generate using Fluke 5522A |

Matrix B
AC Current Measurement

| Range | Frequency | | | | | | | | |
|-------------------|----------------|----------------|----------------|-----------------|----------------|-----------------|----------------|-----------------|------------------|
| | 10 Hz to 20 Hz | 10 Hz to 45 Hz | 20 Hz to 45 Hz | 45 Hz to 100 Hz | 45 Hz to 1 kHz | 100 Hz to 1 kHz | 1 kHz to 5 kHz | 5 kHz to 10 kHz | 10 kHz to 30 kHz |
| 0 to 330 μ A | 28 nA | - | 26 nA | - | 26 nA | - | 33 nA | 72 nA | 340 nA |
| 0.33 mA to 3.3 mA | 630 nA | - | 620 nA | - | 620 nA | - | 630 nA | 650 nA | 920 nA |
| 3.3 mA to 33 mA | 4.9 μ A | - | 3.4 μ A | - | 5.7 μ A | - | 9.9 μ A | 10 μ A | 13 μ A |
| 33 mA to 330 mA | 46 μ A | - | 28 μ A | - | 27 μ A | - | 31 μ A | 36 μ A | 50 μ A |
| 330 mA to 1.1 A | - | 590 μ A | - | - | 160 μ A | - | 200 μ A | 380 μ A | - |
| 1.1 A to 3 A | - | 330 μ A | - | - | 440 μ A | - | 930 μ A | 2 mA | - |
| 3 A to 10 A | - | - | - | 1 mA | - | 1 mA | 4.5 mA | - | - |
| 10 A to 21 A | | | | 6 mA | | 6 mA | 11 mA | | |
| 20 A to 50 A | | | | 300 mA | | | | | |

NO: SAMM 176

Page: 4 of 28

SCOPE OF CALIBRATION: ELECTRICAL

| Instrument Calibrated / Measurement Parameter | Range | Calibration and Measurement Capability Expressed as an Uncertainty (\pm)* | Remarks |
|---|---|---|--|
| 2. Insulation Tester (Max. Voltage 1000 V) | 10 k Ω to 100 k Ω 100 k Ω to 10 M Ω 10 M Ω to 100 M Ω 100 M Ω to 1000 M Ω 1000 M Ω to 10 G Ω | 0.58 Ω 580 Ω 6 k Ω 69 k Ω 600 k Ω | Generate using high resistance decade box IET HRRS |
| 3. Sourcing/Generating Instrument (a) DC Voltage | 0 mV to 100 mV 100 mV to 1 V 1 V to 10 V 10 V to 100 V 100 V to 1000 V | 0.9 μ V 8.8 μ V 58 μ V 0.6 mV 6.1 mV | Measure using System Multimeter HP 3458A |
| (b) DC Current | 0 μ A to 10 mA 10 mA to 100 mA 100 mA to 1 A | 580 nA 7.8 μ A 590 μ A | Measure using System Multimeter HP 3458A |
| (c) AC Voltage | 100 mV to 10 V 40 Hz to 20 kHz 50 kHz to 1 MHz 10 V to 100 V 40 Hz to 1 kHz 100 V to 700 V 40 Hz to 1 kHz | 590 μ V 604 μ V 6 mV 59 mV | Measure using System Multimeter HP 3458A |
| (d) AC Current At 1 kHz | 0 to 10 mA 10 mA to 100 mA 100 mA to 1 A | 71 μ A 23 μ A 780 μ A | Measure using System Multimeter HP 3458A |
| (e) Resistance | 0 Ω to 10 Ω 10 Ω to 100 Ω 100 Ω to 1 k Ω 1 k Ω to 10 k Ω 10 k Ω to 100 k Ω 100 k Ω to 1 M Ω 1 M Ω to 10 M Ω | 665 $\mu\Omega$ 6 m Ω 16 m Ω 150 m Ω 1.7 Ω 24 Ω 730 Ω | Measure using System Multimeter HP 3458A |

Scan this QR Code or visit www.ism.gov.my/cab-directories for the current scope of accreditation

NO: SAMM 176

Page: 5 of 28

SCOPE OF CALIBRATION: ELECTRICAL

| Instrument Calibrated / Measurement Parameter | Range | Calibration and Measurement Capability Expressed as an Uncertainty (\pm)* | Remarks | |
|---|---------------------------------|---|---|---|
| (e) Resistance (continued) | 0 Ω to 10 Ω | 230 $\mu\Omega$ | Measure using System Multimeter HP 3458A | |
| | 10 Ω to 100 Ω | 1.9 m Ω | | |
| | 100 Ω to 1 k Ω | 15 m Ω | | |
| | 1 k Ω to 10 k Ω | 125 m Ω | Measure (direct measurement for resistance boxes) | |
| | 10 k Ω to 100 k Ω | 620 m Ω | | |
| | 100 k Ω to 1 M Ω | 19 Ω | | |
| | 1 M Ω to 10 M Ω | 230 Ω | | |
| | 10 M Ω to 100 M Ω | 13 k Ω | | |
| | | 0 Ω to 1 Ω | 79 $\mu\Omega$ | Measure using System Multimeter HP 3458A Measure (Standard Resistors) |
| | | 1 Ω to 10 Ω | 230 $\mu\Omega$ | |
| 10 Ω to 10 k Ω | | 120 m Ω | | |
| i) DC Current: | 1 A to 2 A | 210 μ A | Measure using System Multimeter HP 3458 A & AC / DC Current Shunt Ballantine 1625 A | |
| | 2 A to 10 A | 3 mA | | |
| ii) AC Current to 1 kHz: | 0 A to 2 A | 1 mA | | |
| | 2 A to 10 A | 21 mA | | |
| iii) AC Current 1 to 5 kHz: | 0 A to 10 A | 22 mA | | |
| iv) AC Current 5 to 10 kHz: | 0 A to 2 A | 1 mA | | |
| | | | | |
| DC Current (Source) | 10 A to 20 A | 3 mA | | |
| | 20 A to 100 A | 3 mA | | |
| AC Current @ 50 Hz to 1 kHz (Source) | 0 A to 20 A | 21 mA | | |
| | 0 A to 100 A | 40 mA | | |
| (f) Frequency | 3 GHz to 5 GHz | 210 nHz | Measuring using Universal Counter Agilent 53132A | |
| | 0 MHz to 3 GHz | 175 nHz | | |

Scan this QR Code or visit www.ism.gov.my/cab-directories for the current scope of accreditation

SCOPE OF CALIBRATION: ELECTRICAL

| Instrument Calibrated / Measurement Parameter | Range | Calibration and Measurement Capability Expressed as an Uncertainty (\pm) * | Remarks |
|---|---|--|--|
| 4. High Voltage Meters/ Testers | DC V: 0 - 40 kV AC V: 0 - 28 kV (peak ac at 50 Hz) | 31 V 120 V | Measure with probes using Digital Multimeter & High Voltage Probe Fluke 80K-40 |
| | DC V: 0 - 2 kV DCV: 2 – 20 kV AC V: 0 - 2 kV at 50 Hz | 600 mV 1.0 V 700 mV | Measure direct using Precision High Voltage Meter Vitrek 4700 |
| | AC V: 2 - 20 kV at 50 Hz | 17 V | |
| | DC V: 0 - 10 kV AC V: 0 - 10 kV at 50 Hz | 1.0 V 17 V | Generate using Withstanding Voltage Tester Kikusui TOS 5101 & Precision High Voltage Meter Vitrek 4700 |
| 5. Clamp Meters | DC Current 0 A to \pm 10 A (up to 1000 A via multi-turn coil – 10 & 50 Turn Coil) | 57 mA | Generate using Calibrator Wavetek 9100 |
| | AC Current 0 A to \pm 10 A 40 Hz to 440 Hz (up to 1000 A via multi-turn coil – 10 & 50 Turn Coil) | 61 mA | |
| | AC Current 10 A to 20 A 40 Hz to 110 Hz | 73 mA | |

NO: SAMM 176

Page: 7 of 28

SCOPE OF CALIBRATION: ELECTRICAL

| Instrument Calibrated / Measurement Parameter | Range | Calibration and Measurement Capability Expressed as an Uncertainty (\pm) * | Remarks | |
|--|-----------------------------------|--|--|---|
| 6. Withstanding / Insulation Voltage Testers (a) Generate: (direct) | DC V: 0 kV to 20 kV | 1.2 V | Measure using Precision High Voltage Meter Vitrek 4700 | |
| | AC V: 0 kV to 20 kV at 50 Hz | 17 V | | |
| | (b) Resistance (Measure) | 10 k Ω to 100 k Ω | | 0.58 Ω |
| | 100 k Ω to 10 M Ω | 577 Ω | | |
| | 10 M Ω to 100 M Ω | 6 k Ω | Generate using high resistance decade box IET HRRS | |
| | 100 M Ω to 1000 M Ω | 69 k Ω | | |
| | 1000 M Ω to 10 G Ω | 600 k Ω | | |
| (c) Cut Off Current (AC Current) | 0 mA to 1.0 mA | 12 μ A | | Measure using Cut Off Current Kikusui TLC 501 B |
| | 1.0 mA to 2.0 mA | 31 μ A | | |
| | 2.0 mA to 5.0 mA | 60 μ A | | |
| | 5.0 mA to 10.0 mA | 110 μ A | | |
| (d) Timing | 0 sec to 60 sec | 62 msec | Measure using Time Calibrator SST-2 | |
| | 1 min to 15 min | 85 msec | | |
| | 15 min to 25 min | 800 msec | | |

Note : dBm is measured relative to 1 mW

SCOPE OF CALIBRATION: ELECTRICAL

| Instrument Calibrated / Measurement Parameter | Range | Calibration and Measurement Capability Expressed as an Uncertainty (\pm) * | Remarks | |
|---|----------------------|--|--|--|
| 7. Oscilloscope | (a) Amplitude | 100 V | 51 mV | Generate using Calibration Generator Tektronix PG 506A |
| | | 50 V | 51 mV | |
| 20 V | | 170 μ V | | |
| 10 V | | 45 μ V | | |
| 5 V | | 23 μ V | | |
| 2 V | | 9.4 μ V | | |
| 1 V | | 4.8 μ V | | |
| 0.5 V | | 2.5 μ V | | |
| 0.2 V | | 1.2 μ V | | |
| 0.1 V | | 900 nV | | |
| 10 mV | | 900 nV | | |
| 1 mV | 900 nV | | | |
| (b) Sweep | 0.1 μ s | 570 ps | Generate using Time Mark Generator Tektronix TG 501A | |
| | 0.2 μ s | 1.2 ns | | |
| | 0.5 μ s | 2.8 ns | | |
| | 1 μ s | 4 ns | | |
| | 2 μ s | 9 ns | | |
| | 5 μ s | 22 ns | | |
| | 10 μ s to 0.1 ms | 460 ns | | |
| | 0.2 ms | 820 ns | | |
| | 0.5 ms | 2 μ s | | |
| | 1 ms | 4.5 μ s | | |
| | 2 ms | 8 μ s | | |
| | 5 ms | 20 μ s | | |
| | 10 ms | 45 μ s | | |
| | 20 ms | 80 μ s | | |
| | 50 ms | 200 μ s | | |
| | 0.1 s | 460 μ s | | |
| | 0.2 s | 930 μ s | | |
| 0.5 s | 2 ms | | | |
| 1 s to 2 s | 4.5 ms | | | |
| 5 s | 23 ms | | | |
| (c) Bandwidth | 0.25 MHz to 250 MHz | 23 kHz | Generate using Levelled Sine Wave Generator Tegam SG 503 Generate using Signal Generator Agilent E4433B | |
| | 250 MHz to 1 GHz | 230 kHz | | |

NO: SAMM 176

Page: 9 of 28

SCOPE OF CALIBRATION: ELECTRICAL

| Instrument Calibrated / Measurement Parameter | Range | Calibration and Measurement Capability Expressed as an Uncertainty (\pm) * | Remarks | | |
|--|---|--|---|-------------------------------|--|
| 8. LCR Meters. Measuring Instruments. i) Inductance: | Frequency at 1 kHz | | Generate using Decade Inductor GR 1491G | | |
| | 100 μ H to 1000 μ H | 110 nH | | | |
| | 1 mH to 10 mH | 800 nH | | | |
| | 10 mH to 100 mH | 12 μ H | | | |
| | 100 mH to 1000 mH | 82 μ H | | | |
| | 1 H to 10 H | 790 μ H | | | |
| | ii) Capacitance: | Frequency at 1 kHz | | | Generate using Decade Capacitor GR 1413 |
| | | 1 pF to 100 pF | | 0.084 pF | |
| | | 100 pF to 10000 pF | | 0.57 pF | |
| | | 10 nF to 100 nF | | 12 pF | |
| 100 nF to 1000 nF | | 570 pF | | | |
| 9. Time | 0 sec to 10 sec | 120 msec | Measure using Time Calibrator SST-2 | | |
| | 0 sec to 100 sec | 120 msec | | | |
| | 0 sec to 1000 sec | 130 msec | | | |
| | 0 sec to 10000 sec | 590 msec | | | |
| | 0 min to 100 min | 590 msec | | | |
| | 0 hr to 24 hr | 590 msec | | | |
| 10. RPM Related Measuring Instruments (Non Contact Type) | 60 rpm to 5999 rpm | 0.07 rpm | In-House Method ESF/0303 | | |
| | 6000 rpm to 29999 rpm | 7 rpm | | | |
| | 30000 rpm to 59999 rpm | 36 rpm | | | |
| | 60000 rpm to 99999 rpm | 120 rpm | | | |
| | | | | | |
| 11. Power Meters/ Indicating Instruments a) DC Power | 0.1 W to 1 kW | $0.26 \frac{mW}{W} + 0.46 \text{ mW}$ | Generate using Fluke 5522A | | |
| | 1 kW to 20 kW | $0.93 \frac{mW}{W} - 120 \text{ mW}$ | | | |
| | b) AC Power 45 Hz to 65 Hz at PF = 1 | 0.1 W to 1 W | $0.59 \frac{mW}{W} + 0.47 \text{ mW}$ | Generate using Fluke 5522A | |
| | | 1 W to 10 kW | $1.1 \frac{mW}{W} + 430 \text{ mW}$ | | |
| | | 10 kW to 20 kW | $1.2 \frac{mW}{W} + 580 \text{ mW}$ | | |
| | | | | | |

Inoperative

Signatories:

- Gurdeep Singh a/l Gurdial Singh
- Mohd Saharudin Bin Rahidi (DC and Low Frequency)

NO: SAMM 176

Page: 10 of 28

SCOPE OF CALIBRATION: ELECTRICAL**SITE: CATEGORY I**

| Instrument calibrated/ Measurement parameter | Range | Calibration and Measurement Capability expressed as an uncertainty (\pm) * | Remarks |
|---|--|---|--|
| 1. Measuring Instruments (a) DC Voltage | 0 mV to 330 mV 330 mV to 3.3 V 3.3 V to 33 V 33 V to 330 V 330 V to 1020 V | 1.4 μ V 15 μ V 160 μ V 880 μ V 8.6 mV | Generation using calibrator model Fluke 5522 A |
| (b) AC Voltage | See Matrix C See Matrix D | See Matrix C See Matrix D | Generation using calibrator model Fluke 5522 A & Wavetek 9100 |

Matrix C
AC Voltage Measurement

| Range | Frequency | | | | | | | | | |
|--------------------|----------------------|----------------------|----------------------|-----------------------|-----------------------|-----------------------|------------------------|------------------------|-------------------------|--------------------------|
| | 10 Hz to 45 Hz | 45 Hz to 1 kHz | 1 kHz to 5 kHz | 5 kHz to 10 kHz | 45 Hz to 10 kHz | 1 kHz to 10 kHz | 10 kHz to 20 kHz | 20 kHz to 50 kHz | 50 kHz to 100 kHz | 100 kHz to 500 kHz |
| 0 mV to 33 mV | 8 μ V | - | - | - | 7.1 μ V | - | 7.2 μ V | 8.1 μ V | 18 μ V | 67 μ V |
| 33 mV to 330 mV | 9.6 μ V | - | - | - | 9.4 μ V | - | 9 μ V | 9.6 μ V | 38 μ V | 83 μ V |
| 330 mV to 3.3 V | 400 μ V | - | - | - | 240 μ V | - | 290 μ V | 400 μ V | 950 μ V | 3.5 mV |
| 3.3 V to 33 V | 1.9 mV | - | - | - | 1.3 mV | - | 1.6 mV | 2.0 mV | 5.3 mV | - |
| 33 V to 330 V | - | 9.6 mV | - | - | - | 15 mV | 17 mV | 18 mV | 130 mV | - |
| 330 V to 1020 V | - | 130 mV | 110 mV | 130 mV | - | - | - | - | - | - |

Scan this QR Code or visit www.ism.gov.my/cab-directories for the current scope of accreditation

NO: SAMM 176

Page: 11 of 28

SCOPE OF CALIBRATION: ELECTRICAL

SITE: CATEGORY I

| Instrument calibrated/ Measurement parameter | Range | Calibration and Measurement Capability expressed as an uncertainty (\pm) * | Remarks |
|---|---|---|--|
| (c) DC Current | 0 to 330 μ A 330 μ A to 3.3 mA 3.3 mA to 33 mA 33 mA to 330 mA 330 mA to 1.1 A 1.1 A to 3 A 3 A to 11 A 11 A to 21 A | 25 nA 170 nA 670 nA 6.7 μ A 280 μ A 530 μ A 2.4 mA 15 mA | Generation using calibrator model Fluke 5522 A |
| (d) AC Current | See Matrix E See Matrix F | See Matrix E See Matrix F | Generation using calibrator model Fluke 5522 A & Wavetek 9100 |

Matrix D
AC Voltage Measurement

| Range | Frequency | | | | | | | | | |
|--------------------|----------------------|----------------------|----------------------|-----------------------|-----------------------|-----------------------|------------------------|------------------------|----------------------------|--------------------------|
| | 10 Hz to 45 Hz | 45 Hz to 1 kHz | 1 kHz to 5 kHz | 5 kHz to 10 kHz | 45 Hz to 10 kHz | 1 kHz to 10 kHz | 10 kHz to 20 kHz | 20 kHz to 50 kHz | 50 kHz to 100 kHz | 100 kHz to 500 kHz |
| 800 V to 1000 V | - | - | - | - | - | - | 1.7 V | - | - | - |

Matrix E
AC Current Measurement

| Range | Frequency | | | | | | | | | |
|-----------------------------|----------------------|----------------------|----------------------|-----------------------|-----------------------|----------------------|----------------------|-----------------------|------------------------|--|
| | 10 Hz to 20 Hz | 10 Hz to 45 Hz | 20 Hz to 45 Hz | 45 Hz to 100 Hz | 100 Hz to 1 kHz | 45 Hz to 1 kHz | 1 kHz to 5 kHz | 5 kHz to 10 kHz | 10 kHz to 30 kHz | |
| 0 μ A to 330 μ A | 120 nA | - | 120 nA | - | - | 120 nA | 180 nA | 240 nA | 480 nA | |
| 0.33 mA to 3.3 mA | 2.6 μ A | - | 1.7 μ A | - | - | 1.5 μ A | 2.6 μ A | 6.2 μ A | 12 μ A | |
| 3.3 mA to 33 mA | 4.4 μ A | - | 3.4 μ A | - | - | 7.0 μ A | 12 μ A | 27 μ A | 51 μ A | |
| 33 mA to 330 mA | 85 μ A | - | 54 μ A | - | - | 37 μ A | 92 μ A | 190 μ A | 370 μ A | |
| 330 mA to 1.1 A | - | 2.2 mA | - | - | - | 0.3 mA | 8.1 mA | 350 mA | - | |
| 1.1 A to 3 A | - | 2.2 mA | - | - | - | 0.81 mA | 8.1 mA | 35 mA | - | |
| 3 A to 11 A | - | - | - | 4.4 mA | 5.8 mA | - | 110 mA | - | - | |
| 11 A to 20.5 A | - | - | - | 21 mA | 24 mA | - | 350 mA | - | - | |

Scan this QR Code or visit www.ism.gov.my/cab-directories for the current scope of accreditation

NO: SAMM 176

Page: 12 of 28

SCOPE OF CALIBRATION: ELECTRICAL**SITE: CATEGORY I**

Matrix F

AC Current Measurement

| Range | Frequency | | | | | | | | |
|--------------|----------------------|----------------------|----------------------|-----------------------|-----------------------|----------------------|----------------------|-----------------------|------------------------|
| | 10 Hz to 20 Hz | 10 Hz to 45 Hz | 20 Hz to 45 Hz | 45 Hz to 100 Hz | 100 Hz to 1 kHz | 45 Hz to 1 kHz | 1 kHz to 5 kHz | 5 kHz to 10 kHz | 10 kHz to 30 kHz |
| 11 A to 20 A | - | - | - | - | - | - | - | 160 mA | - |

| Instrument calibrated/ Measurement parameter | Range | Calibration and Measurement Capability expressed as an uncertainty (\pm) * | Remarks |
|--|---|--|---|
| (e) Resistance | 0 to 11 Ω | 1.2 m Ω | Generation using calibrator model Fluke 5522 A |
| | 11 Ω to 33 Ω | 2.1 m Ω | |
| | 33 Ω to 110 Ω | 2.7 m Ω | |
| | 110 Ω to 330 Ω | 5.9 m Ω | |
| | 330 k Ω to 1.1 k Ω | 35 m Ω | |
| | 1.1 k Ω to 3.3 k Ω | 62 m Ω | |
| | 3.3 k Ω to 11 k Ω | 130 m Ω | |
| | 11 k Ω to 33 k Ω | 620 m Ω | |
| | 33 k Ω to 110 k Ω | 1.3 Ω | |
| | 110 k Ω to 330 k Ω | 6.4 Ω | |
| | 330 k Ω to 1.1 M Ω | 39 Ω | |
| | 1.1 M Ω to 3.3 M Ω | 120 Ω | |
| | 3.3 M Ω to 11 M Ω | 570 Ω | |
| | 11 M Ω to 33 M Ω | 6.4 k Ω | |
| | 33 M Ω to 110 M Ω | 23 k Ω | |
| 110 M Ω to 330 M Ω | 500 k Ω | | |
| 330 M Ω to 1100 M Ω | 6.3 M Ω | | |
| Resistance Up to 1000 Vmax | 100 Ω to 1 k Ω | 140 m Ω | Generate using High Resistance Decade Substitute IET HRRS & Yokogawa 2793 |
| | 1 k Ω to 10 k Ω | 57 Ω | |
| | 10 k Ω to 100 k Ω | 58 Ω | |
| | 100 k Ω to 1000 k Ω | 130 Ω | |
| | 1000 k Ω to 10 M Ω | 1.5 k Ω | |
| | 10 M Ω to 100 M Ω | 24 k Ω | |
| | 100 M Ω to 1000 M Ω 1000 M Ω to 10 G Ω | 270 k Ω 130 M Ω | |
| 0 Ω to 1 Ω | 6 $\mu\Omega$ | ESD/0360 Rev 5.0 Generating using Standard Resistor & Decade Resistance Box model Yokogawa 2792 & 2793 respectively | |
| 1 Ω to 100 Ω | 1 m Ω | | |

Scan this QR Code or visit www.ism.gov.my/cab-directories for the current scope of accreditation

NO: SAMM 176

Page: 13 of 28

SCOPE OF CALIBRATION: ELECTRICAL**SITE: CATEGORY I**

| Instrument calibrated/ Measurement parameter | Range | Calibration and Measurement Capability expressed as an uncertainty (\pm) * | Remarks |
|--|--|--|---|
| (f) Frequency | 0.01 Hz to 120 Hz 120 Hz to 1200 Hz 1.2 kHz to 12 kHz 12 kHz to 120 kHz 120 kHz to 1200 kHz 1.2 MHz to 3.2 MHz 3.2 MHz to 10 MHz | 9.0 μ Hz 9.0 mHz 29 mHz 290 mHz 820 mHz 110 Hz 350 kHz | Generation using calibrator model Fluke 5522 A |
| (g) Capacitance | 220 pF to 400 pF 0.4 nF to 1.1 nF 1.1 nF to 3.3 nF 3.3 nF to 11 nF 11 nF to 33 nF 33 nF to 110 nF 110 nF to 330 nF 0.33 μ F to 1.1 μ F 1.1 μ F to 3.3 μ F 3.3 μ F to 11 μ F 11 μ F to 33 μ F 33 μ F to 110 μ F 110 μ F to 330 μ F 0.33 mF to 1.1 mF 1.1 mF to 3.3 mF 3.3 mF to 11 mF 11 mF to 33 mF 33 mF to 110 mF | 13 pF 17 pF 23 pF 40 pF 98 pF 300 pF 900 pF 4.1 nF 9.3 nF 21 nF 85 nF 290 nF 920 nF 6.4 μ F 19 μ F 38 μ F 130 μ F 540 μ F | Generation using calibrator model Fluke 5522 A |
| Capacitance Frequency at 1 KHz | 1 pF to 100 pF 100 pF to 10000 pF 10 nF to 100 nF 100 nF to 1000 nF | 0.67 pF 1 pF 12 pF 570 pF | ESD/0360 Rev 5.0 Generation using Decade Capacitor model GR 1413 |
| Inductance | 100 μ H to 1000 μ H 1 mH to 10 mH 10 mH to 100 mH 100 mH to 1000 mH 1 H to 10 H | 0.16 μ H 0.28 mH 0.34 mH 0.95 mH 0.3 H | ESD/0360 Rev 5.0 Generating using Decade Inductor model GR 1491-G |

Scan this QR Code or visit www.ism.gov.my/cab-directories for the current scope of accreditation

NO: SAMM 176

Page: 14 of 28

SCOPE OF CALIBRATION: ELECTRICAL**SITE: CATEGORY I**

| Instrument Calibrated/ Measurement Parameter | Range | Calibration and Measurement Capability Expressed as an Uncertainty (\pm) * | Remarks | |
|--|-----------------------------------|---|-------------|--|
| 2. Source / Generating Instruments | (a) DC Voltage | 0 mV to 100 mV | 2.2 μ V | Measured using Multimeter HP 3458A |
| | | 100 mV to 1 V | 14 μ V | |
| | | 1 V to 10 V | 230 μ V | |
| | | 10 V to 100 V | 1.5 mV | |
| | (b) DC Current | 100 V to 1000 V | 15 mV | |
| | | 0 mA to 10 mA | 9.1 μ A | |
| | | 10 mA to 100 mA | 72 μ A | |
| | | 100 mA to 1000 mA | 1.6 mA | |
| | | 1 A to 3 A | 5.7 mA | |
| | | 3 A to 20 A | 6.0 mA | |
| | (c) AC Voltage | 20 A to 100 A | 20 mA | |
| | | 0 mV to 10 mV (40 Hz to 1 kHz) | 5.4 μ V | |
| | | 100 mV to 10 V (40 Hz to 1 kHz) | 1.2 mV | |
| | | 10 V to 100 V (40 Hz to 1 kHz) | 29 mV | |
| | (d) AC Current | 100 V to 1000 V (40 Hz to 1 kHz) | 550 mV | |
| | | 0 A to 1 A (10 Hz to 5 kHz) | 2.3 mA | |
| 1 A to 3 A (10 Hz to 5 kHz) | | 8.1 mA | | |
| 3 A to 20 A (10 Hz to 1 kHz) | | 26 mA | | |
| | 20 A to 100 A (10 Hz to 1 kHz) | 130 mA | | |

Scan this QR Code or visit www.ism.gov.my/cab-directories for the current scope of accreditation

NO: SAMM 176

Page: 15 of 28

SCOPE OF CALIBRATION: ELECTRICAL**SITE: CATEGORY I**

| Instrument Calibrated/ Measurement Parameter | Range | Calibration and Measurement Capability Expressed as an Uncertainty (\pm) * | Remarks | |
|--|---------------------------------|--|--|---------|
| 3. High Voltage Measurement - High Voltage Testers / Puncture Testers / Spark Testers | DC Voltage | | Measured using High Voltage Meter Vitretek 4700A | |
| | | 0 kV to 2 kV | | 1.3 V |
| | | 2 kV to 5 kV | | 7.9 V |
| | | 5 kV to 10 kV | | 11 V |
| | | 10 kV to 15 kV | | 13 V |
| | | 15 kV to 20 kV | | 16 V |
| | AC Voltage | | | |
| | 20 Hz to 100 Hz | 0 kV to 2 kV | | 4 V |
| | | 2 kV to 5 kV | | 30 V |
| | | 5 kV to 10 kV | | 34 V |
| | | 10 kV to 15 kV | | 45 V |
| | | 15 kV to 20 kV | | 57 V |
| | Cut Off Current | | | |
| | AC Current | | | |
| | 40 Hz to 1 kHz | 0 mA to 0.5 mA | | 0.03 mA |
| | 0.5 mA to 1 mA | 0.05 mA | | |
| | 1 mA to 2 mA | 0.06 mA | | |
| | 2 mA to 5 mA | 0.12 mA | | |
| | 5 mA to 10 mA | 0.20 mA | | |
| 4. Timer | 0 sec to 10 sec | 120 msec | Measure using Time Calibrator SST-2 | |
| | 0 sec to 100 sec | 120 msec | | |
| | 0 sec to 1000 sec | 130 msec | | |
| | 0 sec to 10000 sec | 570 msec | | |
| | 0 min to 100 min | 570 msec | | |
| | 0 hr to 24 hr | 570 msec | | |
| 5. Source Instruments Resistance | 0 to 10 Ω | 240 $\mu\Omega$ | Measure using Multimeter HP 3458 A | |
| | 10 Ω to 100 Ω | 2 m Ω | | |
| | 100 Ω to 1 k Ω | 2.9 m Ω | | |
| | 1 k Ω to 10 k Ω | 130 m Ω | | |
| | 10 k Ω to 100 k Ω | 180 m Ω | | |
| | 100 k Ω to 1 M Ω | 20 Ω | | |
| | 1 M Ω to 10 M Ω | 690 Ω | | |
| | 10 M Ω to 100 M Ω | 59 k Ω | | |
| | 100 M Ω to 1 G Ω | 590 k Ω | | |

Scan this QR Code or visit www.ism.gov.my/cab-directories for the current scope of accreditation

NO: SAMM 176

Page: 16 of 28

SCOPE OF CALIBRATION: ELECTRICAL**SITE: CATEGORY I**

| Instrument Calibrated/ Measurement Parameter | Range | Calibration and Measurement Capability Expressed as an Uncertainty (\pm) * | Remarks |
|--|---|--|--|
| 6. Oscilloscope (a) Vertical Deflection (Square Wave Signal) | 100 V 50 V 20 V 10 V 5 V 2 V 1 V 0.5 V 0.2 V 0.1 V 10 mV 1 mV | 51 mV 51 mV 170 μ V 45 μ V 23 μ V 9.4 μ V 4.8 μ V 2.5 μ V 1.2 μ V 900 nV 900 nV 5.2 μ V | Generate using Calibration Generator Tektronix PG 506A |
| (b) Vertical Deflection (DC Signal) | 0mV to 330 mV 330 mV to 3.3 V 3.3 V to 33 V 33 V to 330 V 330 V to 1020 V | 1.6 μ V 23 μ V 24 μ V 290 μ V 3.3 mV | Generation using Calibrator model Fluke 5522 A |
| (C) Sweep | 0.1 μ s 0.2 μ s 0.5 μ s 1 μ s 2 μ s 5 μ s 10 μ s to 0.1 ms 0.2 ms 0.5 ms 1 ms 2 ms 5 ms 10 ms 20 ms 50 ms 0.1 s 0.2 s 0.5 s 1 s to 2 s 5 s | 570 ps 1.2 ns 2.8 ns 4 ns 9 ns 22 ns 460 ns 820 ns 2 μ s 4.6 μ s 8 μ s 20 μ s 45 μ s 80 μ s 200 μ s 460 μ s 930 μ s 2 ms 4.6 ms 23 ms | Generate using Time Mark Generator Tektronix TG 501A |
| (d) Bandwidth | 0.25 MHz to 250 MHz 250 MHz to 4 GHz | 23 kHz 230 kHz | Generate using Levelled Sine Wave Generator Tegam SG 503 & Signal Generator Agilent E4433B |

Scan this QR Code or visit www.ism.gov.my/cab-directories for the current scope of accreditation

NO: SAMM 176

Page: 17 of 28

SCOPE OF CALIBRATION: ELECTRICAL**SITE: CATEGORY I**

| Instrument Calibrated/ Measurement Parameter | Range | Calibration and Measurement Capability Expressed as an Uncertainty (\pm) * | Remarks |
|--|---------------------------------------|---|---|
| 7. Sourcing / Generating Instrument | 0 Hz to 5 GHz | 215 nHz | Measure using Universal Counter Agilent 53132A Measure using Spectrum Analyzer Agilent E4445A |
| (a) Frequency | 5 GHz to 13 GHz | 0.1 μ Hz | |
| (b) Amplitude / Level | -150 dB to +30 dB (3 Hz to 13 GHz) | 1 dBm | |
| 8. Measuring Instrument | 0 MHz to 60 MHz | 5 Hz | Measure using Spectrum Analyzer Agilent E4445A |
| (a) Frequency | 60 MHz to 4 GHz | 1 Hz | Generate using Synthesizer Function Generator HP 3325 B |
| (b) Amplitude / Level | + 20 dBm to -136 dBm | 0.9 dBm | Generate using Signal Generator Agilent E4433B |
| 9. Power Meters/ Indicating Instruments | | | |
| a) DC Power | 0.1 W to 1 kW | $0.26 \frac{mW}{W} + 0.46 \text{ mW}$ | Generate using Fluke 5522A |
| | 1 kW to 20 kW | $0.93 \frac{mW}{W} - 120 \text{ mW}$ | |
| b) AC Power 45 Hz to 65 Hz at PF = 1 | 0.1 W to 1 W | $0.59 \frac{mW}{W} + 0.47 \text{ mW}$ | Generate using Fluke 5522A |
| | 1 W to 10 kW | $1.10 \frac{mW}{W} + 430 \text{ mW}$ | |
| | 10 kW to 20 kW | $1.20 \frac{mW}{W} + 580 \text{ mW}$ | |

Scan this QR Code or visit www.ism.gov.my/cab-directories for the current scope of accreditation**Signatories:**

- Gurdeep Singh a/l Gurdial Singh**
- Mohd Saharudin Bin Rahidi (DC & Low Frequency)**

NO: SAMM 176

Page: 18 of 28

SCOPE OF CALIBRATION: DIMENSIONAL

| Instrument Calibrated/ Measurement Parameter | Range | Calibration and Measurement Capability Expressed as an Uncertainty (\pm) * | Remarks |
|--|---|---|--|
| 1. Micrometer | Up to 25 mm travers 25m to 50 mm 50 mm to 75 mm 75 mm to 100 mm | 1 μ m 1 μ m 2 μ m 2 μ m | Calibrated using Steel Gauge Block as a Standard with reference to ISO 3611: 2010 |
| 2. Caliper | Up to 200 mm 200 mm to 600 mm | 6 μ m 11 μ m | Calibrate using Caliper Checker as a Standard with reference to JIS B 7507:1993 |
| 3. Caliper Checker | 20 mm to 350 mm 350 mm to 450 mm 450 mm to 600 mm | 3 μ m 5 μ m 6 μ m | Calibrate using Steel Gauge Block set as a Standard & Linear Height Gauge as a comparator |
| 4. Standard Rod | 20 mm to 100 mm 100 mm to 150 mm 150 mm to 300 mm 300 mm to 450 mm 450 mm to 800 mm | 2 μ m 3 μ m 4 μ m 5 μ m 10 μ m | Calibrate using Steel Gauge Block set as a Standard & Linear Height Gauge as a comparator |
| 5. Calibration Tester | 0.001 mm to 25 mm | 1.3 μ m | Calibrate using Steel Gauge Block set as a Standard & Sylvac as a Standard |
| 6. Dial Thickness Gauge | 0.001 mm to 50 mm 50 mm to 100 mm | 1 μ m 2 μ m | Calibrate using Gauge Steel Block set as a Standard |
| 7. Pin Gauge (Diameter Only) | 0.2 mm to 40 mm | 2.2 μ m | Using Std. Pin Gauge set as a Standard & Laser Micrometer as a comparator |
| 8. Feeler Gauge | 0.005 mm to 2 mm | 1.0 μ m | Calibrate using Steel Gauge Block set as a Standard & Sylvac Probe as a comparator with reference to JIS-B-7524: 2008 (clause 5:1) |

Scan this QR Code or visit www.ism.gov.my/cab-directories for the current scope of accreditation

NO: SAMM 176

Page: 19 of 28

SCOPE OF CALIBRATION: DIMENSIONAL

| Instrument Calibrated/ Measurement Parameter | Range | Calibration and Measurement Capability Expressed as an Uncertainty (\pm) * | Remarks |
|--|--------------------------------|---|--|
| 9. Dial Test Indicator | 0.01 mm to 1 mm | 3 μ m | Calibrate using Calibration Tester as a Standard with reference to JIS-B-7533:1990 |
| 10. Dial Gauge | 0.01 mm to 100 mm | 4 μ m | Calibrate using Calibration Tester as a Standard with reference to JIS-B 7503:1997 |
| 11. Depth Gauges / Micrometer | 0.001 to 300 mm | 7 μ m | Calibrate using Depth Microchecker as a Standard with reference to BS 6468:2008 or JIS-B 7518:1993 |
| 12. Digital Indicator (Linear Displacement) | 0.001 to 50 mm 50 to 100 mm | 1 μ m 2 μ m | Calibrate using steel gauge block as standard |

Signatories:

1. **Murali a/l Subramaniam**
2. **Mokhtar Ali**

NO: SAMM 176

Page: 20 of 28

SCOPE OF CALIBRATION: DIMENSIONAL**SITE: CATEGORY I**

| Instrument Calibrated/ Measurement Parameter | Range | Calibration and Measurement Capability Expressed as an Uncertainty (\pm) * | Remarks |
|---|------------------------------------|---|--|
| Profile Projector / Tool Maker Microscope / Smart Scope (Measuring accuracies of respective X and Y Axis Only) | up to 80 mm 81 mm to 300 mm | 3 μ m 5 μ m | Calibrate using Glass Scale as a Standard according to JIS B 7153:1995, Performance No. 9 |

Signatories:

1. **Murali a/l Subramaniam**
2. **Mokhtar Ali**

NO: SAMM 176

Page: 21 of 28

SCOPE OF CALIBRATION: TEMPERATURE

| Instrument Calibrated/ Measurement Parameter | Range | Calibration and Measurement Capability Expressed as an Uncertainty (\pm) * | Remarks |
|--|--|--|--|
| 1. Liquid In Glass Thermometer (Total Immersion) | -30 °C to 100 °C 100 °C to 250 °C | 0.08 °C 0.09 °C | Comparison Method Using SPRT Module + PT100 |
| 2. Digital Thermometer with Probe | -40 °C to 250 °C 250 °C to 400 °C 400 °C to 600 °C | 0.10 °C 0.15 °C 0.2 °C | Comparison Method Using SPRT Module + PT100 |
| 3. Mechanical Thermometer | -30 °C to 150 °C 150 °C to 250 °C | 0.3 °C 0.6 °C | Comparison Method Using SPRT Module + PT100 |
| 4. Temperature Indicating Instruments | <u>Type-K</u> -100 °C to 1300 °C <u>Type-J</u> -100 °C to 1200 °C <u>Type-T</u> -100 °C to 400 °C <u>Type-E</u> -100 °C to 1000 °C <u>Type-R</u> 0 °C to 1700 °C <u>Type-S</u> 0 °C to 1700 °C <u>PT100</u> -100 °C to 800 °C | 0.3 °C 0.3 °C 0.3 °C 0.3 °C 0.4 °C 0.4 °C 0.2 °C | Electrical Simulation Using Documenting Process Calibrator based on EA-10/11 and ITS 90 Tables |

Scan this QR Code or visit www.ism.gov.my/cab-directories for the current scope of accreditation

NO: SAMM 176

Page: 22 of 28

SCOPE OF CALIBRATION: TEMPERATURE

| Instrument Calibrated/ Measurement Parameter | Range | Calibration and Measurement Capability Expressed as an Uncertainty (\pm) * | Remarks |
|--|---|---|---|
| 5. Temperature Block Calibrator | -40 °C to 0 °C 0 °C to 600 °C | 0.1 °C 0.2 °C | Using SPRT Module with PT 100 and PRT Based on EA-10/13 |
| 6. Thermo-Hygrometer / Thermo-Hygrograph | 30 %rh to 90 %rh (@ 23 °C) 10 °C to 30 °C 30 °C to 50 °C | 2.5 %rh 0.5 °C 0.6 °C | Comparison method using Std. Temperature and Humidity Meter based on BS 1339- 3:2004 |
| 7. Infrared Thermometer | 35 °C to 100 °C 100 °C to 200 °C 200 °C to 400 °C | 0.7 °C 1.5 °C 2.0 °C | Calibration by Comparison Method using; 1. Infrared Thermometer 2. Blackbody calibrator |

Signatory:

- Zaidi Bin Borham**

NO: SAMM 176

Page: 23 of 28

SCOPE OF CALIBRATION: TEMPERATURE

SITE CALIBRATION: CATEGORY I

| Instrument Calibrated/ Measurement Parameter | Range | Calibration and Measurement Capability Expressed as an Uncertainty (\pm) * | Remarks |
|--|-------------------------------------|---|---|
| 1. Temperature Indicating Instruments | <u>Type-K</u> -100 °C to 1300 °C | 0.4 °C | Electrical Simulation Using Documenting Process Calibrator based on EA-10/11 and ITS 90 Tables |
| | <u>Type-J</u> -100 °C to 1200 °C | 0.4 °C | |
| | <u>Type-T</u> -100 °C to 400 °C | 0.4 °C | |
| | <u>Type-E</u> -100 °C to 1000 °C | 0.4 °C | |
| | <u>Type-R</u> 0 °C to 1700 °C | 0.5 °C | |
| | <u>Type-S</u> 0 °C to 1700 °C | 0.5 °C | |
| | <u>PT100</u> -100 °C to 800 °C | 0.3 °C | |
| 2. Temperature Controlled Enclosure | -40 °C to 250 °C | 0.7 °C | Using Temperature Recorder and Thermocouple Wires based on TLAS G-20 |
| 3. Temperature Sensor with Indicator | -40 °C to 250 °C | 0.3 °C | Calibration by comparison method using: 1. PRT 2. Dry Block Calibrator |
| | 250 °C to 600 °C | 0.4 °C | |

Signatory:

1. Zaidi Bin Borham

NO: SAMM 176

Page: 24 of 28

SCOPE OF CALIBRATION: MASS

| Instrument Calibrated/ Measurement Parameter | Range | Calibration and Measurement Capability Expressed as an Uncertainty (\pm) * | Remarks |
|--|--------------|---|---|
| Standard Weights | Up to 200 mg | 0.018 mg | Comparison Method using ABBA weighing scheme Using E2 Standard Weights by ABBA comparison weighing scheme |
| | 500 mg | 0.020 mg | |
| | 1 g | 0.023 mg | |
| | 2 g | 0.026 mg | |
| | 5 g | 0.032 mg | |
| | 10 g | 0.038 mg | |
| | 20 g | 0.046 mg | |
| | 50 g | 0.06 mg | |
| | 100 g | 0.10 mg | |
| | 200 g | 0.22 mg | |
| | 500 g | 2.9 mg | |
| 1000 g | 3.4 mg | | |
| Standard Weights | 2 kg | 0.11 g | Using Standard Weights F1, F2 ABBA Method |
| | 5 kg | 0.14 g | |
| | 10 kg | 0.17 g | |
| | 20 kg | 0.34 g | |
| | 25 kg | 0.40 g | |

Signatories:

1. **Murali a/l Subramaniam**
2. **Zaidi Borham**

NO: SAMM 176

Page: 25 of 28

SCOPE OF CALIBRATION: MASS**SITE CALIBRATION: CATEGORY I**

| Instrument Calibrated/ Measurement Parameter | Range | Calibration and Measurement Capability Expressed as an Uncertainty (\pm) * | Remarks |
|--|--|---|--|
| 1. Analytical Balance | Up to 200 g | 0.2 mg | Using Std. Weight E2 as a Standard with reference to OIML R 76-I Edition 2006(E) |
| 2. Electronic Balance | Up to 500 g Up to 1 kg Up to 2 kg Up to 5 kg Up to 15 kg Up to 30 kg Up to 60 kg Up to 100 kg Up to 200 kg Up to 500 kg Up to 1000 kg Up to 2000 kg | 1.8 mg 3 mg 6 mg 18 mg 70 mg 2 g 5 g 12 g 19 g 104 g 180 g 204 g | Using Standard Weight E2, F1, F2, M2 as a Standard with reference to OIML R 76-I Edition 2006(E) |

Signatories:

1. **Murali a/l Subramaniam**
2. **Zaidi Borham**

NO: SAMM 176

Page: 26 of 28

SCOPE OF CALIBRATION: FORCE

| Instrument Calibrated/ Measurement Parameter | Range | Calibration and Measurement Capability Expressed as an Uncertainty (\pm) * | Remarks |
|--|--|---|------------------------------------|
| Digital Force Gauges / Push Pull Gauges / Tension Gauges | 1 mgf to 10 kgf 10 kgf to 50 kgf 50 kgf to 100 kgf | 0.6 gf 10 gf 61 gf | Using Std. Weight as a Standard |

Signatory:

1. **Murali a/l Subramaniam**

SCOPE OF CALIBRATION: FORCE**SITE CALIBRATION: CATEGORY I**

| Instrument Calibrated/ Measurement Parameter | Range | Calibration and Measurement Capability Expressed as an Uncertainty (\pm) * | Remarks |
|--|--|---|--|
| Universal Testing Machine (Compression Mode) | 150 kgf to 1000 kgf 1001 kgf to 5000 kgf 5001 kgf to 10000 kgf | 0.2 kgf 3.2 kgf 2.6 kgf | Using Std. Load Cell * Calibration Loop as a Standard with reference to ISO 7500- 1:2004 |

Signatory:

1. **Murali a/l Subramaniam**

NO: SAMM 176

Page: 27 of 28

SCOPE OF CALIBRATION: TORQUE

| Instrument Calibrated/ Measurement Parameter | Range | Calibration and Measurement Capability Expressed as an Uncertainty(\pm)* | Remarks |
|---|------------------|--|--------------------------------------|
| Torque Wrench | 0 Nm to 100 Nm | 0.12 Nm | Measuring using Torque Transducer |
| | 0 Nm to 500 Nm | 0.58 Nm | |
| | 501 Nm to 800 Nm | 1.5 Nm | |

Signatories:

1. **Murali a/l Subramaniam**
2. **Mokhtar Ali**

NO: SAMM 176

Page: 28 of 28

SCOPE OF CALIBRATION: PRESSURE

| Instrument Calibrated/ Measurement Parameter | Range | Calibration and Measurement Capability Expressed as an Uncertainty (\pm) * | Remarks |
|--|------------------|---|---|
| Pressure Indicating Instruments (hydraulic medium) | 0 - 16000 psi | 0.01 % of Reading | Using Dead Weight Tester / Pressure Comparator as a Standard |
| Pressure Indicating Instruments (Pneumatic) | -1 bar to 0 bar | 0.002 bar | Generation Using Digital Pressure Calibrator |
| | 0 bar to 120 bar | 0.01 % of reading | Generating using Air Dead Weight Tester |

Signatories:

1. Murali a/l Subramaniam
2. Mokhtar Ali

SCOPE OF CALIBRATION: PRESSURE**SITE CALIBRATION: CATEGORY I**

| Instrument Calibrated/ Measurement Parameter | Range | Calibration and Measurement Capability Expressed as an Uncertainty (\pm) * | Remarks |
|--|----------------------|---|--|
| Pressure Indicating Instruments (hydraulic medium) | 0 psi to 300 psi | 0.02 % of Reading | Using Pressure Comparator as a Standard |
| | 301 psi to 16000 psi | 0.04 % of Reading | |
| Pressure Indicating Instruments (Pneumatic) | -1 bar to 0 bar | 0.002 bar | Generation Using Digital Pressure Calibrator |
| | 0 bar to 0.07 bar | 0.9% of Reading | |
| | 0.08 bar to 20 bar | 0.01% of Reading | |

Signatories:

1. Murali a/l Subramaniam
2. Mokhtar Ali